

TENNESSEE BUREAU OF INVESTIGATION
Breath Alcohol Unit

Certification Worksheet

Customer Name: _____
Customer Address: _____

Certification occurred at customer address:

Certification occurred at Laboratory:

Date of Certification: _____
Instrument Name: _____
Instrument Serial Number: _____
Procedure Used: _____

Reference Barometer Serial Number: _____ Reference Barometer (mmHg): _____
Reference Barometer Calibration Due: _____ INSTRUMENT Barometer (mmHg): _____
Barometer agreement: _____

Standards (sea level): 0.040g/210L, 0.082g/210L, 0.200g/210L

	0.040g/210L	0.082g/210L	0.200g/210L
Lot Number			
Tank Number			
Expiration			
Target Value			
Sample 1			
Sample 2			
Sample 3			

The calibration of the above instrument has a +/-0.007g/210L measurement certainty at a 99.73% confidence level (k=3)

Passed Certification:
Adjustment:
0.100g/210L Target value: _____
Adjustment tank lot #/ tank #: _____

Replace dry gas standard:
Dry gas lot number/ tank number: _____
Dry gas expiration date: _____

Run a Breath Test (Blank Test):

Download database to IntoxNet:

Sample 1 (g/210L): _____

Sample 2 (g/210L): _____

Notes:

TBI Special Agent Forensic Scientist _____ Date _____